

WYAMZ Kindersley Station - March 2015 Summary Report

*All data has been slope, intercept, and baseline corrected. Data may change after validation process.


*All compliance parameters were within the Saskatchewan Ambient Air Quality Standards.


*All compliance parameters are above 90% operational, there are no operational issues to report.

*The airpointer was calibrated on March 23.




Month Records 24-Hour Records **1-Hour Records Pollutants** Maximum SAAQS Exceed Maximum SAAQS Exceed Conc. Avg. Name Uptime Unit Conc. Standards Conc. Standards Conc. Time Days Time Hours NO 1.1 100.0% 3.3 Mar-24 19.8 Mar-24 08:00 ppb NO₂ 100.0% 6.0 Mar-4 Mar-24 08:00 212 0 ppb 2.9 17.7 -NOx 100.0% Mar-24 08:00 ppb 3.8 8.4 Mar-24 37.0 -_ SO_2 100.0% 4.2 Mar-7 57 0 46.6 Mar-24 17:00 172 0 ppb 1.1






*All compliance parameters are above 90% operational; there are no operational issues to report.




WTAMZ Malusione Station - March 2015 Summary hepott											
Pollutants		Month Records		24-Hour Records				1-Hour Records			
Name	Conc.	Avg.	Uptime	Maximum		SAAQS	Exceed		Maximum		Exceed
Name	Unit	Conc.		Conc.	Time	Objective	Days	Conc.	Time	Objective	Hours
NO	ppb	1.1	82.8%	2.4	Mar-3	-	-	17.9	Mar-10 12:00	-	-
NO ₂	ppb	4.6	82.8%	7.4	Mar-5	-	-	17.3	Mar-05 08:00	212	0
NOx	ppb	5.7	82.8%	8.9	Mar-4	-	-	28.2	Mar-10 12:00	-	-
SO ₂	ppb	0.3	82.8%	0.6	Mar-5	57	0	1.6	Mar-25 15:00	172	0
H ₂ S	ppb	0.2	82.5%	0.5	Mar-18	3.6	0	2	Mar-16 08:00	10.8	0
PM _{2.5}	ug/m3	5	83.9%	12	Mar-25	30	0	29	Mar-25 04:00	-	-

WYAMZ Maidstone Station - March 2015 Summary Report

SAAQS: Saskatchewan Ambient Air Quality Standards







WYAMZ Maidstone Station - March 2015 Summary Report

Monthly Update

*All data has been slope, intercept, and baseline corrected. Data may change after validation process.

*All compliance parameters were within the Saskatchewan Ambient Air Quality Standards.

*No compliance parameters were above 90% operational this month, due to power disruptions and a blown fuse. The airpointer was not


operational March 1 from 14:00 to 22:00, and March 20 01:00 through March 24 18:00.

*The airpointer was calibrated on March 27.



WYAMZ Meadow Lake Station - March 2015 Summary Report											
Pollutants		Month Records		24-Hour Records				1-Hour Records			
Name	Conc.	Avg.	Uptime	Maximum		SAAQS	Exceed	Maximum		SAAQS	Exceed
Name	Unit	Conc.	Optime	Conc.	Time	Objective	Days	Conc.	Time	Objective	Hours
NO	ppb	0.1	100.0%	0.4	Mar-5	-	-	1.4	Mar-05 12:00	-	-
NO ₂	ppb	0.7	100.0%	1.6	Mar-5	-	-	3.5	Mar-27 13:00	212	0
NOx	ppb	0.8	100.0%	2.0	Mar-5	-	-	4.5	Mar-27 13:00	-	-
O ₃	ppb	37	100.0%	44	Mar-8	-	-	48	Mar-08 15:00	82	0
PM _{2.5}	ug/m3	3	100.0%	9	Mar-26	30	0	42	Mar-17 04:00	-	-





*All data has been slope, intercept, and baseline corrected. Data may change after validation process.


*All compliance parameters were within the Saskatchewan Ambient Air Quality Standards.

*All compliance parameters were greater than 90% operational, there are no operational issues to prevent.

*The airpointer was calibrated on March 26.




WYAMZ Meadow Lake Station - March 2015 Summary Report





WYAMZ Unity Station - March 2015 Summary Report											
Pollutants		Month Records		24-Hour Records				1-Hour Records			
Name	Conc.	Avg.	Uptime	Maximum		SAAQS	Exceed	Maximum		SAAQS	Exceed
Indiffe	Unit	Conc.	Optime	Conc.	Time	Objective	Days	Conc.	Time	Objective	Hours
NO	ppb	0.6	87.3%	3.5	Mar-8	-	-	22.3	Mar-08 19:00	-	-
NO ₂	ppb	2.2	87.3%	5.6	Mar-13	-	-	13.6	Mar-08 19:00	212	0
NOx	ppb	2.7	87.3%	7.6	Mar-8	-	-	35.9	Mar-08 19:00	-	-
O ₃	ppb	34	100.0%	45	Mar-30	-	-	57	Mar-30 15:00	82	0
PM _{2.5}	ug/m3	4	100.0%	11	Mar-25	30	0	21	Mar-19 04:00	-	-





Monthly Update

*All data has been slope, intercept, and baseline corrected. Data may change after validation process.

*All compliance parameters were within the Saskatchewan Ambient Air Quality Standards.

*All compliance parameters are above 90% operational, with the exception of the NO/NO2/NOx analyzer.

*The NO/NO2/NOx analyzer was malfunctioning from Feb 26 18:00 MST to March 4 17:00 MST, due to an electronics malfunction.

The airpointer was reset and the analyzer returned to normal operations.

*The airpointer was calibrated on March 25.





